



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/854,560	05/15/2001	Hans Berger	66376-252-7	8137
25269	7590	05/18/2005	EXAMINER	
DYKEMA GOSSETT PLLC FRANKLIN SQUARE, THIRD FLOOR WEST 1300 I STREET, NW WASHINGTON, DC 20005			LANEAU, RONALD	
			ART UNIT	PAPER NUMBER
			3627	

DATE MAILED: 05/18/2005

Please find below and/or attached an Office communication concerning this application or proceeding.

## Office Action Summary

Application No.

09/854,560

Applicant(s)

HANS BERGER ET AL

Examiner

Ronald Laneau

Art Unit

3627

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --  
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

### Status

- 1) ☒ Responsive to communication(s) filed on 16 March 2005.  
2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.  
3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

### Disposition of Claims

- 4) ☒ Claim(s) 1-30 is/are pending in the application.  
4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.  
5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.  
6) ☒ Claim(s) 1-30 is/are rejected.  
7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.  
8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

### Application Papers

- 9) ☐ The specification is objected to by the Examiner.  
10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).  
11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).  
a) ☐ All b) ☐ Some \* c) ☐ None of:  
1. ☐ Certified copies of the priority documents have been received.  
2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.  
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

### Attachment(s)

- 1) ☐ Notice of References Cited (PTO-892)  
2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)  
3) ☐ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)  
Paper No(s)/Mail Date \_\_\_\_\_.  
4) ☐ Interview Summary (PTO-413)  
Paper No(s)/Mail Date \_\_\_\_\_.  
5) ☐ Notice of Informal Patent Application (PTO-152)  
6) ☐ Other: \_\_\_\_\_.

***Response to Amendment***

1. The amendment filed on 3/16/05 has been entered. Claims 1-30 are still pending.

***Specification***

2. Applicant is reminded of the proper content of an abstract of the disclosure.

The abstract should be in narrative form and generally limited to a single paragraph on a separate sheet within the range of 50 to 150 words. It is important that the abstract not exceed 150 words in length since the space provided for the abstract on the computer tape used by the printer is limited. The form and legal phraseology often used in patent claims, such as "means" and "said," should be avoided. The abstract should describe the disclosure sufficiently to assist readers in deciding whether there is a need for consulting the full patent text for details.

The language should be clear and concise and should not repeat information given in the title. It should avoid using phrases which can be implied, such as, "The disclosure concerns," "The disclosure defined by this invention," "The disclosure describes," etc.

***Claim Rejections - 35 USC § 102***

3. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

4. Claims 1-4, 19 and 20 are rejected under 35 U.S.C. 102(b) as being anticipated by Lobiondo et al (US 5,305,199).

Lobiondo et al teach a method for automation of the management of operating materials and/or supplies of an analyzer or analyzing system for determining a parameter or a parameter group of a sample, being used in medical, environmental or food technology, comprising: a) automatic recording of type and maximum useful life of said operating materials used, and of

Art Unit: 3627

types, expiry dates and quantities of said required supplies (col. 3, lines 51-58), b) entering a desired frequency of analysis, or automatic calculation of an estimated frequency of analysis from past frequencies of use of said analyzer or analyzing system (see fig. 3, projected usage), c) automatic calculation of said operating materials and/or supplies required per unit of time, based on data obtained in steps (a) and (b) (col. 3, line 67 to col. 4, line 4), d) determining an optimum point in time for ordering said required operating materials and/or supplies (col. 4, lines 4-9), e) automated ordering of said operating materials and/or supplies via a device for remote data transmission (col. 4, lines 17-23). Furthermore, Lobiondo et al teach a method wherein said operating materials and/or supplies are ordered via an internet connection and wherein said unit for remote data transmission is used to provide an internet portal for information on products, software, service, maintenance, and use, in the fields of medical and food technology (col. 3, lines 16-31).

### ***Claim Rejections - 35 USC § 103***

5. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

6. Claims 5-18 and 21-30 are rejected under 35 U.S.C. 103(a) as being unpatentable over Lobiondo et al (US 5,305,199) in view of Sano et al (US 5,415,840).

Art Unit: 3627

Lobiondo et al further teach updating information on product using a software, expected delivery and shipment information that can be transmitted via a link (col. 3, lines 20-24 and 42-46).

Lobiondo et al do not teach an analyzing system for determining medical sample parameters but Sano et al teach a system for determining medical sample wherein said connection for remote data transmission is provided in a computer central unit of said analyzing system ((fig. 1), the analyzer is coupled to the central unit as claimed and can be removed to be inserted in a different position.(col. 4, lines 42-48), said analyzing system is capable of being provided with a sample bus to exchange the samples to be tested between the analyzer and the control unit (col. 4, lines 1-13).

Neither Lobiondo et al nor Sano et al expressly teach exchanging washing, calibrating and quality control media between analyzer and the control unit but this feature is obvious in the medical field because the system as taught by Sano et al intrinsically would have to do at least some washing, calibrating and quality control media in order to have any kind of exchange between these two elements and also to ensure the reliability of the equipment in use to perform a certain test.

It would have been obvious to one of ordinary skill in the art to utilize the automatic analyzer as taught by Sano et al into the system of Lobiondo et al because it would enable measurement to be performed with simple mechanisms and good reproducibility.

***Response to Arguments***

7. Applicant's arguments filed on 02/07/05 have been fully considered but they are not persuasive.

Applicant argues that there is no hint or any suggestion in LoBiondo for automatic recording maximum useful life of hardware components of the printer or of recording expiry data of any supplies or operating materials. Contrary to applicant's arguments, LoBiondo discloses a recording media that is capable of recording as such. Lobiondo does not explicitly disclose whether the recording is automatic or not but it is well settled that it is not "invention" to broadly provide a mechanical or automatic means to replace manual activity which has accomplished the same result. In re Rundell, 18 CCPA 1290, 48 F. 2d 958, 9 USPQ 220. Applicant's arguments are deemed unpersuasive, Claims 1-30 are finally rejected.

***Conclusion***

8. **THIS ACTION IS MADE FINAL.** Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event,


Art Unit: 3627

however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

9. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Ronald Laneau whose telephone number is (571) 272-6784. The examiner can normally be reached on Mon-Fri from 8:30am - 6:00pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Robert Olszewski can be reached on (571) 272-6788. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.


Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).



Ronald Laneau  
Examiner  
Art Unit 3627

5/16/05

rl

  
Primary Examiner  
5/16/05